



LACS



Low Asbestos Content Scheme Individual Results : Round 018 LACS018

For Laboratory Number : 1640 CRB Analyse Service GmbH

From : 30/01/2023 To : 09/06/2023

Report Issued 22/06/2023

Report No. 25/252/2171

Sample	Your Result	Assigned Result	Score
1	Chrysotile	Chrysotile	0

HSE
Melanie Clunas
 LACS Scheme Co-Ordinator
 Fibres Team : Chemical & Biological Risks Unit

HSE, Science & Research Centre
 Harpur Hill, Buxton, Derbyshire, SK17 9JN - UK

FTPT F0931

ISSUE 2

Page 1 of 1





Group Report



Round 18
June 2023

LOW ASBESTOS CONTENT SCHEME

Round 18 Sample Details

BACKGROUND

This report covers Round 18 of the Low Asbestos Content Scheme (LACS). LACS is currently a qualitative only PT scheme.

Round 18 was open to laboratories worldwide. Laboratory participation was as follows: 4 UK and 120 Rest of Europe.

125 laboratories subscribed to this round with 124 submitting results.

SAMPLES

One sample was circulated as follows: Sample LACS018 – This sample was soil containing UICC chrysotile.

SCREENING & VALIDATOR INFORMATION

The sample was prepared for circulation following our normal internal screening process of samples with representative sub-samples scanned using stereo-zoom and polarised light microscopy and transmission electron microscopy to assess homogeneity and suitability. Approximately 10% of the total number of samples despatched were validated by 11 independent laboratories.

INFORMATION SUBMITTED BY LABORATORIES

Laboratories used the PT Online Data Entry System to submit their results for this round. Results were submitted as asbestos type(s) present.

LACS RESULTS

Sample LACS018

Of the 124 laboratories submitting results for R18:

One hundred and twenty two laboratories correctly reported chrysotile

One laboratory reported tremolite


One laboratory reported no asbestos

These results are presented graphically in Charts 1 and 2.



1. Type Of Errors Obtained


Chart 1 illustrates the errors made by participating laboratories. Three errors were made by laboratories on sample LACS018. One falsely reported tremolite and one reported no asbestos.



False Negative = Component has been missed. False Positive = Component has been incorrectly identified as present.

2. Errors for UK & Non-UK Laboratories

Chart 2 illustrates the distribution of scores for all participating laboratories. 122 (98%) laboratories obtained a score of zero in this round, indicating that these laboratories had not made any errors. The distribution of scores obtained by UK (United Kingdom) and Non-UK laboratories is also compared; 4 (100%) UK laboratories and 118 (98%) Non-UK laboratories obtained a score of zero for the round.



Quantitative Results

For 2023/24, HSE are not providing quantitative samples for the Low Asbestos Content Scheme (LACS).

Additional Information

